

報告號碼 : ECR2338378

頁數(page) : 1 of 3

Report No.



報告日期 : 2023/9/25

Report Date

申請者 : 臺北醫學大學附設醫院

Applicant

設備名稱 : 溫濕度計

Equipment Name

顧客住址 : 110臺北市信義區吳興街252號

Customer Address

量測日期 : 2023/9/23

Measurement Date

量測項目 : 溫濕度量測

Measurement Item


適用規範 : -

Measurement Code

量測結果 : 詳見內頁

Measurement Results

This document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <http://www.sgs.com.tw/Terms-and-Conditions> and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com.tw/Terms-and-Conditions>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested.

	Report No. : ECR2338378	頁數(page) : 2 of 3
---	-------------------------	-------------------

申請者 Applicant	臺北醫學大學附設醫院				
設備名稱 Equipment	溫濕度計				
製造廠商 Manufacturer	虹勝	機型 Model	RCH-1	序號 Serial No.	環境溫濕度
作業程序 Procedure used	GENP-EC-T008(V 1.9)	收件日期 Received Date	2023/9/23	量測日期 Measurement Date	2023/9/23
檢驗者 Examiner	蔡鎮任	溫度 °C Temperature	(24 ~ 25) °C	相對濕度 % Relative Humidity	(67 ~ 65) %
顧客住址 Customer Address	110臺北市信義區吳興街252號				

### 量測使用儀器 / SGS Standards

儀器名稱 Equipment	製造廠商 Manufacturer	機型 Model	標準器校正日期 Calibration Date
Humidity and Temperature Meter	VAISALA	RFL100	2023/2/14
序號 Serial Number	追溯單位 Traceability	報告號碼 Report No.	標準器有效日期 Due Date
sensor1~sensor30(API)	SGS(TAF 0143)	ECR2304576	2024/2/13

- ★ 除非獲得本公司書面同意，報告不得摘錄複製。  
Reproduced report in partial is not effective without approval by SGS.
- ★ 檢測結果僅對被檢樣品有效。  
The results of the inspection are only effective for the samples inspected.
- ★ 附圖或照片如附頁  
The sketch or photograph As refer to following pages.
- ★ 判定結果詳內頁  
Please refer to following pages to see the inspection result.
- ★ 作業程序名稱：溫濕度計(LOGGER)(實測比對法+遊校)校正作業程序書  
Procedure use Calibration SOP for Temperature-hygrometer/Logger(Actual Test + On-Site)

蔡鎮任

檢驗者 (Examiner)

丁中明

簽署人 (Signature)

This document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <http://www.sgs.com.tw/Terms-and-Conditions> and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com.tw/Terms-and-Conditions>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested.



報告號碼(Report No.) : ECR2338378

頁數(page) : 3 of 3

一. 溫濕度量測 :

(一) 量測程序 :

1. 確認空調系統已穩定運作。
2. 溫濕度標準件放置於待測溫濕度計旁，溫濕度穩定後進行記錄。

(二) 量測結果 :

器示值		標準值		器差值	
溫度(°C)	相對濕度(%)	溫度(°C)	相對濕度(%)	溫度(°C)	相對濕度(%)
26.3	55.9	26.4	56.2	-0.1	-0.3

---- THE END ----

壹股份