

校正報告書

校正醫量測實驗室-台北

Report No. : ECR2619527

第1頁 共2頁

申請者 臺北醫學大學附設醫院

申請者	臺北醫學大學附設醫院		
儀器名稱	-80°C超低溫冷凍櫃		
製造廠商	PANASONIC	機型	MDF-C8V1-P
序列號	15100231	校正日期	2026/5/19
校正程序	GENP-EC-T014(V 1.7)	收件日期	2026/5/19
Procedure used	校正程序		
校驗者	蔡鎮任	溫度 °C	(20 ~ 21) °C
Operator	11031臺北市信義區吳興街252號		
顧客地址	11031臺北市信義區吳興街252號		
遊校地址	11031臺北市信義區吳興街252號蔡萬才大樓3樓		
Location	校正地點		
遊校	Relative Humidity		
	相對濕度 % (62 ~ 65) %		
	Calibration Date		
	校正日期 2026/5/19		

實驗室使用標準器 / SGS Standards

儀器名稱	製造廠商	機型	標準器校正日期
Thermometer	CHY	506R	2025/7/31
儀器名稱	製造廠商	機型	標準器校正日期
Equiment	Manufacturer	Model	Calibration Date
序號	追溯單位	報告號碼	標準器有效日期
Serial Number	Traceability	Report No.	Due Date
50000511	SGS(TAF 0143)	ECR2529794	2026/7/30

◆台灣檢校科技股份有限公司特此聲明本報告書內記載之標準器，依ILAC P10之規定，可追溯至ILAC-MRA國際實驗室認證聯盟互承認協議成員之認可實驗室，或各國家計量標準機構(NMI)，或國際度量衡委員會(TAF)認證範圍之校正項目證書；無TAF認證標誌之報告亦符合本實驗室標準校正作業程序及ISO/IEC 17025之規定。有TAF認證標誌之報告表示符合全國認證基金會(TAF)認證範圍之校正項目證書；無TAF認證標誌之報告亦符合本實驗室標準校正作業程序及ISO/IEC 17025之規定。

SGS Taiwan Ltd hereby declare that traceability follows ILAC P10(ILAC policy on the traceability of measurement results) All standards are directly traceable to TAF recognized lab (members of the ILAC-MRA) or to National Metrology Institutes (NMI) or to other international standards (members of the CIPM-MRA) or certified reference materials(CRMs) The report with "TAF" accredited symbol represents the certificate of calibration items conforming to TAF accreditation scope. Without "TAF" accredited symbol, the report also complies with the lab's standard calibration operating procedures and ISO/IEC 17025 requirements

◆本校正報告僅對上述儀器之校正項目有效且未經調整，報告內容不包含校正週期之建議以及符合性判定，除特別聲明外。
 ◊ The calibration report is only valid for the instrument mentioned above and has not been adjusted.
 Unless otherwise specified, the content of the report does not include the suggestion for calibration interval and the determination of compliance.

◆本校正報告部份複製及影本無效。

◊ To reproduce or copy calibration report in partial is not allowed

◆校正程序名稱：烘箱(高低溫恆溫箱或冰箱等)校正作業程序書

◊ Procedure used Calibration SOP for Oven (High and low temperature incubator or refrigerator, etc)

報告簽署人

劉俊廷

This document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <https://www.sgs.com.tw/terms-of-service> and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <https://www.sgs.com.tw/terms-of-service>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested in the condition it was received.



RptNo.: ECR2619527

第2頁 共2頁

溫度部份：(比對校正)

器示值 (°C)	-82.8	標準值 (°C)	-82.7	器差值 (°C)	-0.1
----------	-------	----------	-------	----------	------

校正說明：

1. 器差值=器示值 - 標準值
2. 器示值係指送校正件所顯示或設定之值
3. 標準值係指工作標準件之輸出值或顯示值
4. 校正能力係以約95%信賴水準,k=2之相對擴充不確定度表示
5. 擴充不確定度：0.5 °C
6. 上述擴充不確定度已包含校正件之不確定度評估結果

-- THE END --

This document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <https://www.sgs.com/tw/terms-of-service> and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <https://www.sgs.com/tw/terms-of-service>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested in the condition it was received.