


2024/09/01 00:00-2024/10/01 05:00



2024/9/21 15:57-17:25 癌醫大樓進行斷電測試,臨床試驗藥局藥櫃與溫度監測系統皆不受影響,但因網路訊號斷訊,造成溫濕度數值無法讀取而被記錄成"0"度。第2頁為負80°C的backup溫度紀錄,顯示斷訊當下,溫度不受影響,無超溫。

your arm the

01. Oct. very

藥師徐 儀 君
藥字第026959號

負 80 °C

ID: DF6130645B5D472E

Temperature Unit of Measure is Celsius

Date format is MM/DD/YYYY

Start Time and Date

11:30 09/09/2024

Stop Time and Date

15:00 10/01/2024

Minimum Temperature

P1	-85.42	08:19 09/22/2024
P2	---,--	--:-- --/--/----

Maximum Temperature

P1	-77.43	15:11 09/10/2024
P2	---,--	--:-- --/--/----

Alarm Event Data

Probe	Temp	Time Out	Date Out	Time In	Date In
-------	------	----------	----------	---------	---------

Temperature Data

P1	P2	Time	Date
-84.21	---,--	17:45	09/21/24
-85.11	---,--	17:30	09/21/24
-83.93	---,--	17:15	09/21/24
-83.93	---,--	17:00	09/21/24
-84.09	---,--	16:45	09/21/24
-85.05	---,--	16:30	09/21/24
-84.14	---,--	16:15	09/21/24
-83.52	---,--	16:00	09/21/24
-84.34	---,--	15:45	09/21/24
-83.71	---,--	15:30	09/21/24
-84.2	---,--	15:15	09/21/24
-85.18	---,--	15:00	09/21/24
-84	---,--	14:45	09/21/24
-83.79	---,--	14:30	09/21/24
-84.62	---,--	14:15	09/21/24
-85.04	---,--	14:00	09/21/24